

LASER SAFETY INSTRUCTION 2023

FRITZ-HABER-INSTITUT
MAX-PLANCK-GESELLSCHAFT



Instructor: Daniel Wegkamp
wegkamp@fhi-berlin.mpg.de





FORMAL FOUNDATIONS & REGULATIONS

EU Directive 2006/25/EC on the minimum health and safety requirements regarding the exposure of workers to risks arising from **artificial optical radiation**

Germany: Occupational Safety and Health Ordinance on **Artificial Optical Radiation (OStrV)**

Applied regulations: Technical Rules regarding the OStrV (TROS Laser)



Duties of employers:

- Instructions of employees
- Documentation of the instructions
- Annual training
- Operating instructions for the Labs/Lasers (Betriebsanweisung)

Responsible people for laser safety in the department of Physical Chemistry:


Daniel Wegkamp (wegkamp@fhi-berlin.mpg.de)

Marcel Krenz (krenz@fhi-berlin.mpg.de)



OUTLINE

General Laser Safety Instruction (Focus on typical laser systems in the department).
Dangers specific to certain Experiments: refer to annual specific safety instruction within work groups.



OUTLINE

1. Hazards due to laser radiation
2. Laser labs, safety measures & rules of conduct
3. In case of an accident?



1. HAZARDS DUE TO LASER RADIATION


- What is special about Laser Radiation (safety aspects)
- Laser Classes
- Dangers of Laser Radiation



LASER RADIATION


...is collimated

Common Lightsource



Distributed Radiation

LASER = Light Amplification by Stimulated Emission of Radiation




Confined Beam → Continuous High Intensity



LASER RADIATION

...is broadband

➤ wavelength range: 100 nm - 1 mm



→ large part is invisible, but still harmful



LASER CLASSES

| Class | Basis for Classification |
|---|---|
| Class 1 inherently Safe (VIS/non-VIS) | Lasers which are safe under reasonably foreseeable conditions of operation. |
| Class 1 Laser Product (safe if not modified) | Product that contains a higher class laser system but access to the beam is controlled by engineering means. |
| Class 2 Low Power (VIS only) | Protection of the eye is normally provided by natural aversion blink response* which takes ca. 0.25s. These lasers are not intrinsically safe. AEL = 1mW for CW laser. |
| Class 1M Safe without viewing aids (302.5 nm - 4000 nm) | Safe under reasonable foreseeable conditions of operation. Beams are either highly divergent or collimated but with a large diameter. May be hazardous if user employs optics with the beam. |
| Class 2M Safe without viewing aids (VIS only) | Protection of the eye is normally provided by natural aversion blink response* which takes ca. 0.25s. Beams are either highly divergent or collimated but with a large diameter. May be hazardous if user employs optics with the beam. |
| Class 3R Low/medium power (VIS/non-VIS) | Risk of injury is greater than for lower classes but not as high as for class 3B. Up to 5 times the AEL for Class 1 and Class 2. |
| Class 3B Medium/high power (VIS/non-VIS) | Direct intrabeam viewing of these devices is always hazardous. Viewing diffuse reflections is normally safe provided the eyes is no closer than 13 cm from the diffuse surface and the exposure duration is less than 10 seconds. AEL = 500mW for CW laser. |
| Class 4 High power (VIS/non-VIS) | Direct intrabeam viewing is dangerous. Specular and diffuse reflections are hazardous. Eye, skin and fire hazard. TREAT CLASS 4 WITH CAUTION. |



LASER CLASSES

Examples

Class 1



Class 2



Class 3






LASER CLASSES

Examples

Class 4



Direct intrabeam viewing is dangerous. Specular and diffuse reflections are hazardous. Eye, skin and fire hazard. TREAT CLASS 4 WITH CAUTION.

Manufacturer has to classify and mark their product!

Einfaches Erkennungsmerkmal für die Gefährlichkeit von Lasern




Beispiele für die Kennzeichnung von Lasern DIN EN 60825-1:2008-05



DANGERS OF LASER RADIATION

Overview: eye and skin damage






DANGERS OF LASER RADIATION

Eye damage

- Collimated beam is **focused** on retina
- Type of damage depends on wavelength
- Aversion blink response reflex* only works for visible light




* aversion blink response reflex only present in 25% of all people and rarely fast *enough*



DANGERS OF LASER RADIATION

Eye damage in numbers




$$I_{\text{Laser}} = 2.1 \times 10^9 \times I_{\text{Lightbulb}}$$




DANGERS OF LASER RADIATION

Permanent eye damage

Immediately after accident



Only 5% of normal vision after 3 years



Strahlquelle: Rubinlaser (694nm); Pulsenergie: 20mJ;
Pulsdauer: 20ns; Pulsleistung: 1MW; Entfernung: 25m;
Sehstärke: 5% nach Unfall, 5% nach 3 Jahren; Persönliche
Empfindung: schwarzer Fleck, abnehmende Sehschärfe

Retinal damage is permanent!



DANGERS OF LASER RADIATION

Visual effects of eye damage

Damage to the eye is **permanent** and results in loss of vision!



No Damage



Damaged Cornea




Damaged Retina (Blind Spot)



DANGERS OF LASER RADIATION

Skin damage

Thermal Skin Injuries



UV (or ultrashort NIR) laser can cause skin cancer (accumulative effect)






DANGERS OF LASER RADIATION

Common Radiation Pathways


...direct radiation

e.g. laser beam directly into the eye




...reflected radiation

e.g. on wristwatches or jewellery



...diffuse reflection

e.g. on tools or edges



Protection from direct beam is limited **even with laser safety goggles**

Even detector cards scatter and potentially reflect harmful levels of laser radiation





SECONDARY HAZARDS...



- Electrical shock
- Fire & explosion hazard
- Laser generated air contaminants (LGACs)
- **Short laser pulses: nonlinear effects generate additional wavelengths**
- Chemical hazards: Laser dyes and solvents

Methanol
(Index-Nr.: 603-001-00-X)




Gefahr

Gefahrenhinweise
H225 Flüssig und fest brennend
H301 Giftig
+ H311
+ H331
H370 Schädlich für die Umwelt

Sicherheitsmaßnahmen
P210 Von Hitze, offener Flamme, heißen Oberflächen fernhalten
P280 Schutzhandschuhe, Schutzkleidung, Schutzbrille oder Gesichtsschutz
P302 BEI KONTAKT MIT DER HAUT: Alle kontaminierten Kleidungsstücke sofort ausziehen. Haut mit Wasser abwaschen/befeuchten.
+ P332 wuschern
P307 Bei Exposition: RUFEN SIE SOFORT AN EINEM GUT BELÜFTETEN ORT
+ P311 Arzt anrufen.
P403 Behälter dicht verschlossen an einem gut belüfteten Ort aufbewahren.
+ P233 aufbewahren.



IR → UV





2. LASER LABS, SAFETY MEASURES & RULES OF CONDUCT

- Laser labs in the dept. of PC
- Laser safety equipment
- Behaviour in the lab
- Laser safety goggles



LASER LABS & SAFETY MEASURES

General rules

- Laboratories in the department of Physical Chemistry are suitable for operations of class 4 lasers and are marked as such.
- Only instructed personnel is allowed to enter, when lasers are in operation (warning signs).
- Personal safety equipment has to be used (laser safety goggles).
- Laser radiation has to be confined to the laser table(s) by 20 cm tall shielding.



LASER LABS & SAFETY MEASURES

Warning signs

➤ Check laser warning signs *before* you enter the lab!



No Danger Laser Off



Attention Laser On

Laser operators / scientists are responsible for correct display switch setting
Personal safety equipment (laser goggles) has to be used

➤ Only instructed personnel is allowed to enter the lab when lasers are on



LASER LABS & SAFETY MEASURES

Nominal ocular hazard area

Definition:

Nominal ocular hazard area (NOHA)
„Lasergefahrenbereich“



Area in which hazardous laser radiation can be present


Normal Operation

NOHA is confined to the laser table if...

- ... protective shielding is present around the laser table (photo)
- ... laser beam runs parallel to laser table surface
- ... nobody is actively manipulating the beam

Laser beam MUST NOT leave laser table!

Personal safety equipment (laser goggles) has to be used





LASER LABS & SAFETY MEASURES

Nominal ocular hazard area

Definition:

Nominal ocular hazard area (NOHA)
„Lasergefahrenbereich“



Area in which hazardous laser radiation can be present

Maintenance case:

- ... within the laser system itself or
- ... beams leaving the laser table or uncontrolled beams on the table for a limited period

Labs have to be marked with additional warning signs & barriers can be installed



WARNUNG/WARNING
Lasergefahrenbereich ausgeweitet
Laser hazard area extended
Please call lab or contact
_____ before entering.
Always wear laser safety goggles.

Personal safety equipment (laser goggles) has to be used





LASER LABS & SAFETY MEASURES

Inside the Lab

Minimize Risk:

- Workbenches at elevated height
- Elevated seats/chairs (office chairs are prohibited)
- Protect eyes when you have to bend down
- Wear safety goggles even when you are not working with the laser, but someone else is

**Do not compromise the safety of others!
Communicate frequently with your colleagues!**





RULES OF CONDUCT IN THE LASER LAB

Inside the Lab

Optimize work outfit:

- Remove Wristwatches
- Remove Jewelry and other accessories (rings, necklaces, etc.)
- Avoid shiny tools (tweezers, scissors, screwdrivers) also for UHV work



Working at the setup:

- Block beam before inserting or moving optics
- Avoid loose optics!
- Use suitable beam dumps (watercooled)
- Use tools for adjustment (fluorescent cards, cameras, etc.) **Be aware of reflections!**

Wear safety goggles!







RULES OF CONDUCT IN THE LASER LAB

Inside the Lab

- Use suitable beam dumps
- Use beam blocks carefully
- **Use detector cards carefully**





LASER SAFETY GOGGLES

Laser Goggles are personal protective gear!

Each employee...

- ...gets personalized goggles with suitable specs
- ...has the responsibility to make sure the goggles stay intact
- ...must return the goggles when they leave the institute





LASER SAFETY GOGGLES



Product Specification

YAD


EN207Certified
Luminous transmittance:11% Amber

CE Rating

Optical Density

| Wavelength (nm) | L-Ratings |
|--------------------|------------|
| 180-315 | D LB7 + R |
| >315-534 + 730-740 | D LB5 + IR |
| >740-1070 | D LB6 + IR |

Spectrophotometer Data



Product Specification

YG4

EN207Certified
Luminous transmittance:37% Green

CE Rating

Optical Density

| Wavelength (nm) | ODs |
|--------------------|-----|
| 180-315 | 6+ |
| >315-534 + 730-740 | 6+ |
| >740-1070 | 7+ |



The protective effect only exists
if goggles are suitable
for the wavelength of the laser(s)!



3. IN CASE OF AN ACCIDENT



IN CASE OF AN ACCIDENT

Immediate response

Immediately

Transport to


Hochschulambulanz Charité

Campus Benjamin Franklin – Augenklinik

Hindenburgdamm 30

12203 Berlin-Steglitz

Phone: 030 8445 3015



First Aid



Emergency medical eye service at the hospital rescue centre.
The ambulance is located at the north entrance on the ground floor.
It can be reached by telephone at 030 8445 3015.



IN CASE OF AN ACCIDENT

After immediate care

...afterwards:

- Report accident to group leader
- Complete „accident report“ for employer’s liability insurance association
- Visit a „transit doctor* (Durchgangsarzt)“



| Name: | Address: | | Phone: (030-) |
|---|-------------------------|--------------|------------------|
| Dr. Abed Domah | Hohenzollerndamm 124 | 14199 Berlin | 2522569 |
| Dr. Matthias Decker/Dr. Marco Kiesewetter (Orthopädie) | Teltower Damm 15 | 14169 Berlin | 8113106 |
| Prof. Dr. W. K. Ertel, Charité-Campus Benjamin Franklin | Hindenburgdamm 30 | 12203 Berlin | 84453081 |
| Orthopädiegemeinschaft | Breitenbachplatz 21 | 14195 Berlin | 82007430 |
| Orthopädische Gemeinschaftspraxis | Schloßstr. 111 | 12163 Berlin | 7918087 |
| Christoph Olbrich, Krankenhaus Waldfriede | Argentinische Allee 40 | 14163 Berlin | 818100 |
| Prof. Dr. Wolf Petersen, Martin-Luther-Krankenhaus | Caspar-Theyß-Str. 27-31 | 14193 Berlin | 89553025 |
| Dr. Dietmar Sander, St. Gertrauden Krankenhaus | Paretzer Str. 12 | 10713 Berlin | 82722751 |
| Dr. Sebastian Vahrmeyer (Unfallarzt) | Hildegardstr. 28 | 10715 Berlin | 85771427 |
| Dr. Burkard Franz Wolf | Kirchstraße 2 | 14163 Berlin | 8021034 |
| Dr. Thomas Wojtecki | Albrechtstraße 36a | 12167 Berlin | 7916005 |

The Fritz Haber Institute belongs to the „Verwaltungs-Berufsgenossenschaft“

* special medical doctor who cooperates with the employers' liability insurance association. Current list available online and via administration



FURTHER INFORMATION

& thank you for your attention!

Today's instructor:

Daniel Wegkamp

Responsible people for laser safety in the Department of Physical Chemistry:

| | | | |
|----------------|--------|--------|-----------------------------|
| Daniel Wegkamp | G 0.15 | ☎ 5200 | 💻 wegkamp@fhi-berlin.mpg.de |
| Marcel Krenz | G 2.07 | ☎ 5115 | 💻 krenz@fhi-berlin.mpg.de |

Further reading & summary (.pdf):

<https://pctech.rz-berlin.mpg.de/information/>